


Traceability And Uncertainty: Albumin FS
Product Code 1 0220


Calibrator:	TruCal U	Product Code 5 9100
Analyte:	Albumin	
Method:	Bromcresol green method	
Reference:	ERM-DA470	
Total Uncertainty:	4.08%	
Literature:	1. ISO GUM (1993) Guide to the expression of uncertainty in measurement. ISO, Geneva, Switzerland	
	2. EURACHEM Guide (1995) Quantifying uncertainty in analytical measurement. EURACHEM, Teddington, UK	



2007-12-14
Date


Product / Project Management

校准品：	复合校准品	货号 5 9100
被分析物：	白蛋白	
方法：	溴甲酚绿法	
参考物：	ERM-DA470	
总不确定度：	4.08%	
文献：	1. ISO GUM (1993) Guide to the expression of uncertainty in measurement. ISO, Geneva, Switzerland	
	2. EURACHEM Guide (1995) Quantifying uncertainty in analytical measurement. EURACHEM, Teddington, UK	



2007-12-14
日期


生产 / 项目经理

