

BS EN ISO 11138-2:2017



BSI Standards Publication

Sterilization of health care products — Biological indicators

Part 2: Biological indicators for ethylene
oxide sterilization processes (ISO
11138-2:2017)

National foreword

This British Standard is the UK implementation of EN ISO 11138-2:2017. It supersedes BS EN ISO 11138-2:2009 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization and Associated Equipment and Processes.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2017.
Published by BSI Standards Limited 2017

ISBN 978 0 580 89832 7

ICS 11.080.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2017.

Amendments/corrigenda issued since publication

Date	Text affected
<hr/>	

English Version

**Sterilization of health care products - Biological indicators
- Part 2: Biological indicators for ethylene oxide
sterilization processes (ISO 11138-2:2017)**

Stérilisation des produits de santé - Indicateurs
biologiques - Partie 2: Indicateurs biologiques pour la
stérilisation à l'oxyde d'éthylène (ISO 11138-2:2017)

Sterilisation von Produkten für die
Gesundheitsfürsorge - Biologische Indikatoren - Teil 2:
Biologische Indikatoren für Sterilisationsverfahren mit
Ethylenoxid (ISO 11138-2:2017)

This European Standard was approved by CEN on 19 January 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 11138-2:2017) has been prepared by Technical Committee ISO/TC 198 “Sterilization of health care products in collaboration with Technical Committee CEN/TC 102 “Sterilizers and associated equipment for processing of medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2017 and conflicting national standards shall be withdrawn at the latest by September 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11138-2:2009.

The standard is a full technical revision of the previous version. The following amendments have been made in comparison with EN ISO 11138-2:2009:

- requirements on population and resistance (clause 9) revised, e.g. information to minimum *D*-value at 30 °C deleted;
- Annex A, in particular A.2.4 step 6 revised;
- informative Annex B on rationale for the inclusion of a second *D*-value and deletion of the requirement for a minimum *D*-value at 30 °C added;
- informative Annex ZA respective relationship between this European Standard and the essential requirements of Directive 93/42/EEC [OJ L 169] aimed to be covered was deleted.

EN ISO 11138 consists of the following parts, under the general title *Sterilization of health care products — Biological indicators*:

- *Part 1: General requirements*
- *Part 2: Biological indicators for ethylene oxide sterilization processes*
- *Part 3: Biological indicators for moist heat sterilization processes*
- *Part 4: Biological indicators for dry heat sterilization processes*
- *Part 5: Biological indicators for low-temperature steam and formaldehyde sterilization processes*

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11138-2:2017 has been approved by CEN as EN ISO 11138-2:2017 without any modification.

Contents		Page
Foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	General requirements	1
5	Test organism	1
6	Suspension	2
7	Carrier and primary packaging	2
8	Inoculated carriers and biological indicators	2
9	Population and resistance	2
Annex A (normative) Method for determination of resistance to ethylene oxide sterilization		4
Annex B (informative) Rationale for the inclusion of a second minimum <i>D</i> value specification as a result of changes to the test gas used to evaluate resistance and deletion of the requirement for a minimum <i>D</i> value at 30 °C		5
Bibliography		6

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This third edition cancels and replaces the second edition (ISO 11138-2:2006), which has been technically revised.

A list of all parts of ISO 11138 can be found on the ISO website.

Introduction

ISO 11138-1 specifies production, labelling, test methods and performance requirements for the manufacture of biological indicators including inoculated carriers and suspensions intended for use in validation and monitoring of sterilization processes. This document gives specific requirements for those biological indicators intended for use in ethylene oxide sterilization processes.

The ISO 11138 series represents the current “state-of-the-art” according to the experts representing manufacturers, users and regulatory authorities involved in developing the standard. The intent is not to promote the use of biological indicators where such use is not advised, but rather to provide common requirements for the production of those biological indicators that are known to be in use today.

Standards exist providing requirements for the validation and control of ethylene oxide sterilization (see ISO 11135 and ISO 14937).

NOTE It is possible that some countries or regions have published other standards covering requirements for sterilization or biological indicators.

Advice on selection, use and interpretation of results when using biological indicators can be found in ISO 14161.

Sterilization of health care products — Biological indicators —

Part 2: Biological indicators for ethylene oxide sterilization processes

1 Scope

This document specifies requirements for test organisms, suspensions, inoculated carriers, biological indicators and test methods intended for use in assessing the performance of sterilizers and sterilization processes employing ethylene oxide gas as the sterilizing agent, either as pure ethylene oxide gas or mixtures of this gas with diluent gases, at sterilizing temperatures within the range of 29 °C to 65 °C.

NOTE 1 Requirements for validation and control of ethylene oxide sterilization processes are provided by ISO 11135 and ISO 14937.

NOTE 2 National or regional regulations can provide requirements for work place safety.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11138-1:2017, *Sterilization of health care products — Biological indicators — Part 1: General requirements*

ISO 18472, *Sterilization of health care products — Biological and chemical indicators — Test equipment*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11138-1 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp/>

4 General requirements

The requirements of ISO 11138-1 apply.

5 Test organism

5.1 The test organisms shall be spores of *Bacillus atrophaeus*, *Bacillus subtilis* or other strains of microorganisms of demonstrated equivalent performance as required by this document.

NOTE 1 Some strains of *Bacillus subtilis* have been reclassified as *Bacillus atrophaeus*.

NOTE 2 *Bacillus atrophaeus* ATCC 9372, NCTC 10073, NCIMB 8058, DSM 2277, NRRL B-4418 and CIP 77.18 have been found to be suitable¹⁾.

5.2 If a test organism other than *Bacillus atrophaeus* is used, the suitability of the resistance of that test organism shall be determined.

6 Suspension

The requirements of ISO 11138-1 apply.

7 Carrier and primary packaging

7.1 The suitability of the carrier and primary packaging materials for biological indicators for use in ethylene oxide sterilization processes shall be demonstrated in accordance with the requirements of ISO 11138-1:2017, 5.2 and Annex B.

7.2 The exposure conditions to determine compliance shall be

- a) minimum exposure temperature: greater than or equal to 55 °C,
- b) sterilizing agent: ethylene oxide gas at a concentration not less than 800 mg/l in air at greater than or equal to 70 % RH,
- c) maximum exposure temperature: as stated by the biological indicator manufacturer, and
- d) exposure time: greater than or equal to 6h.

NOTE These conditions have been selected to represent a realistic challenge to the carrier while remaining within the practical limits of an ethylene oxide sterilization process.

8 Inoculated carriers and biological indicators

The requirements of ISO 11138-1 apply.

9 Population and resistance

9.1 The manufacturer shall state the resistance characteristics in accordance with ISO 11138-1:2017, 6.4.

9.2 The viable count shall be stated with increments less than or equal to $0,1 \times 10^n$ per unit (e.g. per ml of suspension, per inoculated carrier or per biological indicator).

9.3 For inoculated carriers and biological indicators, the viable count shall be greater than or equal to $1,0 \times 10^6$.

9.4 The resistance shall be expressed as the *D* value in minutes, at 54 °C. The *D* value of each batch/lot of biological indicators or inoculated carriers shall be stated in minutes, to one decimal place at 54 °C.

9.5 Suspensions, inoculated carriers or biological indicators containing *Bacillus atrophaeus* spores shall have a *D* value of not less than 2,5 min at 54 °C when tested according to the conditions in [Annex A](#) using test gas mixtures (see [Annex B](#)). Other microorganisms shall have *D* values supporting the application.

1) These are examples of suitable products available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of these products.

9.6 Suspensions, inoculated carriers or biological indicators containing *Bacillus atrophaeus* spores shall have a *D* value of not less than 2,0 min at 54 °C when tested according to the conditions in [Annex A](#) using a test gas consisting of 100 % EO (see [Annex B](#)). Other microorganisms shall have *D* values supporting the application.

9.7 The resistance characteristics specified in this document and any other parts of ISO 11138 shall be defined using the specific critical variables associated with the referenced sterilization process.

9.8 *D* values are determined according to methods given in ISO 11138-1:2017, Annex C and Annex D.

9.9 Determination of *D* value and survival-kill response characteristics require the use of a resistometer applying the reference resistometer process parameters (see [Annex A](#)).

9.10 The survival-kill window can be calculated using the formulae in ISO 11138-1:2017, Annex E.

NOTE This information can be of value to the user when comparing different batches from the same manufacturer.

EXAMPLE Using the formulae in ISO 11138-1:2017, Annex E with the minimum population and minimum *D* value requirements specified in this document, the survival-kill response characteristics are:

- at 54 °C, 2,5 *D* value; survival time greater than or equal to 10 min and kill time less than or equal to 25 min;
- at 54 °C, 2,0 *D* value; survival time greater than or equal to 8 min and kill time less than or equal to 20 min.

Annex A **(normative)**

Method for determination of resistance to ethylene oxide sterilization

A.1 Principle

This method requires the use of a test apparatus referred to as a resistometer in this document. The specifications of the resistometer process parameters for ethylene oxide sterilization processes are provided in ISO 18472.

Specific requirements related to the test method are provided in [A.2](#).

A.2 Procedure

A.2.1 Load the samples on to suitable sample holders.

A.2.2 Preheat the resistometer chamber to the test condition of 54 °C.

A.2.3 Place the loaded sample holders in the chamber, close the chamber and initiate the test cycle.

A.2.4 Carry out the following sequence of operations:

- Step 1: Evacuate the chamber to a vacuum set point of 10 kPa \pm 0,5 kPa.
- Step 2: Admit sufficient water vapour to raise the relative humidity in the chamber to 60 % \pm 10 %. Maintain these conditions for a period of 30 min \pm 1 min. The samples should be allowed to warm to above the dew point prior to injection of water vapour to avoid the potential for condensation.
- Step 3: Admit ethylene oxide gas to the chamber to obtain a concentration of 600 mg/l \pm 30 mg/l within 60 s. For the 0 min exposure time, no ethylene oxide gas shall be admitted.
- Step 4: Maintain these conditions for the required exposure time \pm 5 s.
- Step 5: At the end of the exposure time, evacuate the chamber to 10 kPa or less within 60 s and then admit filtered air, or an inert gas (such as nitrogen) to ambient pressure.
- Step 6: Repeat step 5 four additional times or more if necessary to reduce operator exposure to any sterilant remaining in the chamber.
- Step 7: At the end of the above test cycle, remove the samples from the chamber and transfer the samples to the growth medium and incubate (see ISO 11138-1:2017, Clause 7).

A.2.5 The transfer period shall be documented and the same time period shall be used for all tests.

A.3 Determination of resistance

Resistance characteristics shall be determined according to methods given in ISO 11138-1:2017, Annexes C, D and E.

Annex B (informative)

Rationale for the inclusion of a second minimum *D* value specification as a result of changes to the test gas used to evaluate resistance and deletion of the requirement for a minimum *D* value at 30 °C

B.1 Inclusion of a second minimum *D* value

This document requires *D* value testing of ethylene oxide (EO) biological indicators (BIs) in an ISO 18472 compliant resistometer using ethylene oxide gas at a nominal concentration of 600 mg/l. Historically, this was often accomplished by BI manufacturers using test gas mixtures. As of 31 December 2014, the U.S. EPA Clean Air Act prohibited the sale and use of HCFC-based (hydrochlorofluorocarbon) products in the United States, requiring U.S.-based BI manufacturers to change the *D* value test gas from HCFC/EO to 100 % EO.

Round robin *D* value testing at 54 °C completed by three U.S.-based biological indicator manufacturers on four different biological indicator products demonstrated that *D* values obtained using 100 % EO were consistently lower than *D* values obtained in the HCFC/EO mixture, even though the nominal EO concentration for both sets of tests was 600 mg/l. The reported differences between *D* values were as high as 40 %. This study has now been published in *Pharmaceutical Technology* (2014)^[4].

ISO 11138-2:2006 has been amended by changing the 54 °C *D* value requirement to add an option of a 2,0 min *D* value specification for tests performed in 100 % EO, based on data reported in the peer reviewed publication. Providing an option for BI manufacturers to maintain compliance with the standard without changing the actual BI resistance was preferred over driving manufacturers to increase BI resistance to meet the standard. This scenario would create more risk of positive BIs for end users whose sterilization processes were validated against the lower resistance BIs.

B.2 Deletion of the requirement for a *D* value at 30 °C

Due to missing data to support the requirement for a 12,5 minute minimum *D* value at 30°C, and the absence of evidence for the actual use of this test method, it has been deleted from this document.

Bibliography

- [1] ISO 11135, *Sterilization of health-care products — Ethylene oxide — Requirements for the development, validation and routine control of a sterilization process for medical devices*
- [2] ISO 14161, *Sterilization of health care products — Biological indicators — Guidance for the selection, use and interpretation of results*
- [3] ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*
- [4] KRUSHEFSKI G. "Effects of 100% Ethylene Oxide Test Gas on the Resistance of Ethylene Oxide Biological Indicators," *Pharmaceutical Technology* 38 (12) 2014

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email subscriptions@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK



医课汇
公众号
专业医疗器械资讯平台
WECHAT OF
HLONGMED



hlongmed.com
医疗器械咨询服务
MEDICAL DEVICE
CONSULTING
SERVICES



医课培训平台
医疗器械任职培训
WEB TRAINING
CENTER



医械宝
医疗器械知识平台
KNOWLEDG
ECENTEROF
MEDICAL DEVICE



MDCPP.COM
医械云专业平台
KNOWLEDG
ECENTEROF MEDICAL
DEVICE