

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

| VALIDATION REPORT 验证报告  |  |
|---|--|
| Product Family Full Cycles Qualification in current Sterilizer<br>产品族在现有灭菌柜的全周期确认 |  |
| File Number 文件编号  |  |

| Revision History 修订历史 |                         |                        |
|-----------------------|-------------------------|------------------------|
| Version 版本号           | Description 描述          | Written By/起草者 Date/日期 |
| 01                    | 1 <sup>st</sup> Version |                        |

| Validation Team<br>验证团队 | Name/Position<br>姓名/职位 | Signature<br>签名 | Date<br>日期 |
|-------------------------|------------------------|-----------------|------------|
| Written By<br>起草者       |                        |                 |            |
| Reviewer<br>审核者         |                        |                 |            |
| Reviewer<br>审核者         |                        |                 |            |
| Reviewer<br>审核者         |                        |                 |            |
| Approver<br>批准者         |                        |                 |            |

Medical Device Manufacturing Co;Ltd  
医疗器械有限公司

目录

|                                    |    |
|------------------------------------|----|
| 1. 验证目标 Validation Objective ..... | 3  |
| 2. 报告总结 Report Conclusions .....   | 3  |
| 3. 实施概述 Execution Summary .....    | 8  |
| 4. 纠正措施 Corrective Actions: .....  | 17 |
| 5. 可交付性 Deliverables: .....        | 17 |
| 6. 附录 Appendices .....             | 17 |

## 1. 验证目标 Validation Objective

此次全周期确认的目的是证实及用文件记录在现有的灭菌柜内采用的预先设定的正常的灭菌循环，通过全周期循环，与半周期确认的 PPQ 数据合并，完成完整的 PPQ，以体现过程的再现形。同时，通过全周期循环，确认产品 EO/ECH 残留自然解析的最短解析时间。此方案的设计持续符合 ISO 11135:2014-医疗产品灭菌-环氧乙烷-医疗器械灭菌过程的发展，验证及常规控制的要求。

The purpose of this Full Cycle Qualification is to complete the full PPQ combined with the PPQ data in half cycles qualification runs for the sterilization of exiting product family in existing EO Sterilization Chamber through the execution of using predetermined nominal full routine sterilization cycle to demonstrate the reproducibility of the process. Also through product EO/ECH residuals tests to determine the minimum ambient aeration time. This protocol is designed to be consistent with ISO 11135:2014 Sterilization of health care products - Ethylene oxide-Requirements for development, validation and routine control of a sterilization process for medical devices.

## 2. 报告总结 Report Conclusions

2.1.证实 Have demonstrated that:

2.1.1. 所有在外置挑战装置内的生物指示剂及 EPCD，全部失活。

All BIs from the processed EPCDs are all deactivated.

2.2.在现有灭菌柜用于此次全周期确认的产品为模拟最具挑战性的非销售工程样品，适用于以下产品。

The full cycle qualification run was conducted in existing EO sterilization chamber with non-saleable engineering products simulating the most challenging routine load were used , and applicable for the products given below:

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

| 序号<br>No. | 代码<br>Code | 产品描述<br>description  | 产品名称<br>Name  | 外箱尺寸<br>dimension(cm) | 每箱数量<br>Quantity per<br>box(pcs) | 包装重量/箱<br>Weight/box<br>(kg) | 灭菌箱数/层数<br>Boxes/layers | 产品体积<br>Volume(m³) | 密度 density<br>(kg/m³) |
|-----------|------------|--|---|-----------------------|----------------------------------|------------------------------|-------------------------|--------------------|-----------------------|
| 1         | JP-2186    | 硅胶圆形开槽管 10Fr 不带引导针<br>(粘接) JP CHANNEL DRAIN<br>10FR, FULL FLUTES | Silicone round drain, full<br>channels without trocar | 47*36*25              | 80                               | 1.3                          | 60/3                    | 0.042              | 30.952                |
| 2         | JP-2187    | 硅胶圆形开槽管 10Fr 带引导针<br>(粘接) JP CHAN DRN SIL RND<br>10FR FULL W/TRO | Silicone round drain, full<br>channels with trocar    | 47*36*25              | 80                               | 2.0                          | 60/3                    | 0.042              | 47.619                |
| 3         | JP-2188    | 硅胶圆形开槽管 15Fr 不带引导针<br>(粘接) JP CHANNEL DRAIN<br>15FR, FULL FLUTES | Silicone round drain, full<br>channels without trocar | 47*29*47              | 80                               | 1.8                          | 60/3                    | 0.064              | 28.125                |
| 4         | JP-2189    | 硅胶圆形开槽管 15Fr 带引导针<br>(粘接) JP CHAN DRN SIL RND<br>15FR FULL W/TRO | Silicone round drain, full<br>channels with trocar    | 47*29*47              | 80                               | 3.8                          | 60/3                    | 0.064              | 59.375                |
| 5         | JP-2190    | 硅胶圆形开槽管 19Fr 不带引导针<br>(粘接) JP CHANNEL DRAIN<br>19FR, FULL FLUTES | Silicone round drain, full<br>channels without trocar | 47*29*47              | 80                               | 2.1                          | 60/3                    | 0.064              | 32.813                |
| 6         | JP-2191    | 硅胶圆形开槽管 19Fr 带引导针<br>(粘接) JP CHAN DRN SIL RND<br>19FR FULL W/TRO | Silicone round drain, full<br>channels with trocar    | 47*29*47              | 80                               | 4.1                          | 60/3                    | 0.064              | 64.063                |
| 7         | JP-2210    | 硅胶扁平开槽管 7mm 不带引导针<br>3/4JP CHANNEL DRAIN 7MM,<br>3/4 FLUTES      | Silicone flat drain without<br>trocar                 | 47*36*25              | 80                               | 2.1                          | 60/3                    | 0.042              | 50.000                |

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

|    |         |   |   |          |    |     |      |       |        |
|----|---------|---|---|----------|----|-----|------|-------|--------|
| 8  | JP-2211 | 硅胶扁平开槽管 7mm 不带引导针<br>全开槽 JP CHANNEL DRAIN<br>7MM, FULL FLUTES       | Silicone flat drain without<br>trocar             | 47*36*25 | 80 | 1.7 | 60/3 | 0.042 | 40.476 |
| 9  | JP-2212 | 硅胶扁平开槽管 7mm 带引导针<br>全开槽 JP CHAN DRN SIL FLT<br>7MM FULL W/TRO       | Silicone flat drain with<br>trocar                | 47*36*25 | 80 | 3.7 | 60/3 | 0.042 | 88.095 |
| 10 | JP-2213 | 硅胶扁平开槽管 10mm 不带引导<br>针 3/4JP CHANNEL DRAIN<br>10MM, 3/4 FLUTES      | Silicone flat drain without<br>trocar             | 47*36*25 | 80 | 2.3 | 60/3 | 0.042 | 54.762 |
| 11 | JP-2214 | 硅胶扁平开槽管 10mm 不带引导<br>针 全开槽 JP CHANNEL DRAIN<br>10MM, FULL FLUTES    | Silicone flat drain without<br>trocar             | 47*36*25 | 80 | 2.1 | 60/3 | 0.042 | 50.000 |
| 12 | JP-2215 | 硅胶扁平开槽管 10mm 带引导针<br>全开槽 JP CHAN DRN SIL FLT<br>10MM FULL W/TRO     | Silicone flat drain with<br>trocar                | 47*36*25 | 80 | 3.8 | 60/3 | 0.042 | 90.476 |
| 13 | JP-2216 | 硅胶扁平开槽管 7mm 带引导针<br>3/4 JP CHAN DRN SIL FLT 7MM<br>3/4 W/TRO        | Silicone flat drain with<br>trocar                | 47*36*25 | 80 | 3.9 | 60/3 | 0.042 | 92.857 |
| 14 | JP-2217 | 硅胶扁平开槽管 10mm 带引导针<br>3/4 JP CHAN DRN SIL FLT<br>10MM 3/4 W/TRO      | Silicone flat drain with<br>trocar                | 47*36*25 | 80 | 4.0 | 60/3 | 0.042 | 95.238 |
| 15 | JP-2221 | 硅胶圆形开槽管 10Fr 带引导针<br>3/4 (粘接) JP CHAN DRN SIL<br>RND 10FR 3/4 W/TRO | Silicone round drain, 3/4<br>channels with trocar | 47*36*25 | 80 | 2.1 | 60/3 | 0.042 | 50.000 |

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

|    |         |   |   |          |    |     |      |       |        |
|----|---------|---|---|----------|----|-----|------|-------|--------|
| 16 | JP-2223 | 硅胶圆形开槽管 15Fr 带引导针<br>3/4（粘接）JP CHAN DRN SIL<br>RND 15FR 3/4 W/TRO | Silicone round drain, 3/4<br>channels with trocar | 47*29*47 | 80 | 4.0 | 60/3 | 0.064 | 62.500 |
| 17 | JP-2225 | 硅胶圆形开槽管 19Fr 带引导针<br>3/4（粘接）JP CHAN DRN SIL<br>RND 19FR 3/4 W/TRO | Silicone round drain, 3/4<br>channels with trocar | 47*29*47 | 80 | 4.1 | 60/3 | 0.064 | 64.063 |
| 18 | JP-2226 | 硅胶圆形开槽管 10Fr 不带引导针<br>JP CHANNEL DRAIN 10FR<br>HUBLESS            | Hubless silicone round<br>drain without trocar    | 47*36*25 | 80 | 1.2 | 60/3 | 0.042 | 28.571 |
| 19 | JP-2227 | 硅胶圆形开槽管 10Fr 带引导针 JP<br>CHANNEL DRAIN 10FR<br>HUBLESS             | Hubless silicone round<br>drain with trocar       | 47*36*25 | 80 | 2.0 | 60/3 | 0.042 | 47.619 |
| 20 | JP-2228 | 硅胶圆形开槽管 15Fr 不带引导针<br>JP CHANNEL DRAIN 15FR<br>HUBLESS            | Hubless silicone round<br>drain without trocar    | 47*29*47 | 80 | 2.1 | 60/3 | 0.064 | 32.813 |
| 21 | JP-2229 | 硅胶圆形开槽管 15Fr 带引导针 JP<br>CHAN DRN SIL HUBLESS 15FR<br>W/TRO        | Hubless silicone round<br>drain with trocar       | 47*29*47 | 80 | 3.9 | 60/3 | 0.064 | 60.938 |
| 22 | JP-2230 | 硅胶圆形开槽管 19Fr 不带引导针<br>JP CHANNEL DRAIN, 19FR<br>HUBLESS           | Hubless silicone round<br>drain without trocar    | 47*29*47 | 80 | 3.0 | 60/3 | 0.064 | 46.875 |
| 23 | JP-2231 | 硅胶圆形开槽管 19Fr 带引导针 JP<br>CHAN DRN SIL HUBLESS 19FR<br>W/TRO        | Hubless silicone round<br>drain with trocar       | 47*29*47 | 80 | 5.8 | 60/3 | 0.064 | 90.625 |

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

|    |         |  |   |          |    |      |      |       |         |
|----|---------|--|---|----------|----|------|------|-------|---------|
| 24 | JP-2232 | 硅胶圆形开槽管 19Fr 带可弯曲引导针 JP CHAN DRN SIL HUBLES 19FR BND TRO | Hubless silicone round drain with bendable trocar             | 47*29*47 | 80 | 5.6  | 60/3 | 0.064 | 87.500  |
| 25 | JP-2233 | 硅胶圆形开槽管 15Fr 带可弯曲引导针 JP CHAN DRN SIL HUBLES 15FR BND TRO | Hubless silicone round drain with bendable trocar             | 47*29*47 | 80 | 3.7  | 60/3 | 0.064 | 57.813  |
| 26 | JP-2234 | 硅胶圆形开槽管 24Fr 带引导针 JP CHANNEL DRAIN, 24FR HUBLESS         | Hubless silicone round drain with trocar, 43" total length    | 47*29*47 | 80 | 9.1  | 60/3 | 0.064 | 142.188 |
| 27 | JP-2290 | 硅胶圆形开槽管 28Fr 不带引导针 CHANNEL DRAIN, 28FR, HUBLESS          | Hubless silicone round drain without trocar, 43" total length | 47*29*47 | 80 | 5.07 | 60/3 | 0.064 | 79.219  |
| 28 | JP-2292 | 硅胶圆形开槽管 32Fr 不带引导针 CHANNEL DRAIN, 32FR, HUBLESS          | Hubless silicone round drain without trocar, 43" total length | 47*29*47 | 80 | 6.57 | 60/3 | 0.064 | 102.656 |

### 3. 实施概述 Execution Summary

#### 3.1.前提步骤 Prerequisite Steps

3.1.1. 在全周期确认活动实施前，确认现有灭菌柜的安装确认及运行确认报告，相关报告已存档。

Verification of the IQ and OQ documentation of existing chamber had been done prior to the full cycle qualification, the documents were archived.

3.1.2. 相关人员的培训已实施，培训记录见附录 1。

Training of the full cycle qualification run was conducted for all relevant personnel involved and the training record see attachment 1.

3.1.3. 核实温湿度无线探头，安装于灭菌柜上的温湿度压力传感器及其他辅助设施的计量仪表的校验记录是否符合接受要求，计量仪表校验状态见仪表校验清单，符合接受要求。仪表校验状态记录见附录 2。

All calibration records/results of the temperature/ relative humidity data loggers, the temperature,RH sensors & pressure transmitter fitted to the sterilization chamber and other measurement instruments associated with the ancillary equipment were reviewed for acceptance. The calibration status for each equipment was identified in the equipment calibration list and was found to be acceptable. The record of equipment calibration status is presented in attachment 2.

3.1.4. 核对了设备维护的状态。所有的设备均依照灭菌设备维护程序规定的时间表进行维护，并用文件记录维护活动，维护记录存档。

The status of equipment maintenance was also reviewed. All identified equipment was conducted according to the time schedule. The maintenance activities were documented and the records were archived in files.

#### 3.2.验证装载 Validation Load

3.2.1. 完整的非销售产品(满载)用于此次全周期循环确认，产品装载体积 4.1 m<sup>3</sup>。

The complete load of non-saleable products (full load) had been used for this full cycle qualification process. The product load volume as derived in protocol is 4.1 m<sup>3</sup>.

3.2.2. 灭菌柜内产品的装载，外置挑战装置，无线温湿度探头的放置参考验证方案附录 2。



## Medical Device Manufacturing Co;Ltd

### 医疗器械有限公司

The load configuration in the chamber, and the placement of EPCDs, and data loggers (temperature and humidity sensors) in the load were carried out according to attachment 2 of the protocol.

#### 3.3. 过程挑战装置 PCDs

3.3.1. 用于此次全周期确认的为外置挑战装置，其制备参考验证方案要求进行。

EPCDs were used in this full nominal cycle qualification. They were prepared according to the related requirements .

3.3.2. 外置挑战装置枯草芽孢杆菌，平均菌落数为 $\geq 1.0 \times 10^6$ /条，D 值在  $54^\circ\text{C} \geq 2.5$  分钟，符合要求。

The EPCDs were inoculated with biological indicator (*Bacillus atrophaeus*). The average spore concentration was  $\geq 1.0 \times 10^6$ /strip and the D value of the BI was  $\geq 2.5$  minutes at  $54^\circ\text{C}$  which met the acceptance criteria as defined .

3.3.3. 此次全周期确认的半周期所用的外置挑战装置为 9 个。

A total of 9 units of EPCD for this full nominal cycle run were used in this full cycle qualification.

#### 3.4. 装载条件 Load Condition

产品装载在进行半周期循环前，需移入冷库冷冻至少 12h, 模拟最恶劣的条件。

The validation load needs be transferred to the the freezer to storage for at least 12h simulating the worst case condition before half cycle runs.

3.5. 产品装载将植入温湿度无线探头，外置过程挑战装置。

The load will be seeded with temperature and humidity data loggers, EPCDs, all EPCDs and sensors will be placed into the full load shipper cases and the EPCDs will be affixed outside the accordingly shipper cases.

#### 3.6. 无线探头 Data Loggers

3.6.1. 校验完的温湿度无线探头用于全周期确认。

Calibrated temperature and humidity data loggers were used in this full cycle qualification run.

3.6.2. 总计 6 个温度，3 个湿度探头依照验证方案附录 2 的要求放置在装载指定的位置。

A total of 6 T and 3 RH data loggers were placed in the locations stipulated in appendix 2 of the protocol.

### 3.7.结果 Results

#### 3.7.1. 过程参数 Process Parameters

3.7.2. 全周期循环符合规定的过程参数，完成的循环参数报告（验证方案附录 1）和灭菌循环记录包括设定值及范围见本报告附录 3。

The full cycle run met the cycle parameters. The filled cycle parameter sheets (appendix 1 of the protocol) and the cycle run record including set points and tolerance are enclosed within this completion report in attachment 3.

### 3.8. 装载的物理性质 Physical profiles of the loads

#### 3.8.1. 装载的温湿度分布见附录 4

The temperature and humidity profiles are presented in attachment 4.

#### 3.8.2. 装载的温湿度分布在半周期循环不同阶段的表现见如下列表：

A summary of the load response at different phases of the three half cycle runs is tabulated below:

Medical Device Manufacturing Co;Ltd  
医疗器械有限公司

表 1Table 1: 装载温度分布 Load Temperature Data Profile (补全信息, 完成后删除此内容)

| PHASES 阶段                  | Load Parameter 装载参数   | 1 <sup>st</sup> half cycle<br>半周期 1 | Position<br>位置 | 2 <sup>nd</sup> half cycle<br>半周期 2 | Position<br>位置 | 3 <sup>rd</sup> half cycle<br>半周期 3 | Position<br>位置 | full cycle<br>全周期 | Position<br>位置 |
|----------------------------|---|-------------------------------------|----------------|-------------------------------------|----------------|-------------------------------------|----------------|-------------------|----------------|
| Start of the cycle<br>循环开始 | Min. Load Temp.最低装载温度   | 4.6℃                                | T1             | -9.4℃                               | T2             | ℃                                   |                |                   |                |
| Preconditioning<br>预热      | Min. Load Temp.最低装载温度   | 4.6℃                                | T1             | -9.4℃                               | T2             | ℃                                   |                |                   |                |
|                            | Max. Load Temp.最高装载温度   | 43.1℃                               | T6             | 46.8℃                               | T5             | ℃                                   |                |                   |                |
|                            | Min. Load Humidity<br>最低湿度  | 73.7%                               | H3             | 44.5%                               | H3             | %                                   |                |                   |                |
|                            | Max. Load Humidity 最高湿度   | 88.8%                               | H1             | 94.7%                               | H1             | %                                   |                |                   |                |
|                            | Min. Load Temp. at the<br>end of Preconditioning<br>Phase 预热结束后的最低温度  | 29.8℃                               | T1             | 31.2℃                               | T1             | ℃                                   |                |                   |                |
|                            | Max. Load Temp. at the<br>end of Preconditioning<br>Phase 预热结束后的最高温度  | 43.1℃                               | T6             | 46.8℃                               | T5             | ℃                                   |                |                   |                |
|                            | Min. Load Humidity at the<br>end of Preconditioning<br>Phase 预热结束最低湿度 | 73.7%                               | H3             | 44.5%                               | H3             | %                                   |                |                   |                |

Medical Device Manufacturing Co;Ltd

医疗器械有限公司

| PHASES 阶段                  | Load Parameter 装载参数   | 1 <sup>st</sup> half cycle<br>半周期 1 | Position<br>位置 | 2 <sup>nd</sup> half cycle<br>半周期 2 | Position<br>位置 | 3 <sup>rd</sup> half cycle<br>半周期 3 | Position<br>位置 | full cycle<br>全周期 | Position<br>位置 |
|----------------------------|---|-------------------------------------|----------------|-------------------------------------|----------------|-------------------------------------|----------------|-------------------|----------------|
|                            | Max. Load Humidity at the end of Preconditioning phase 预热结束最高湿度 | 88.1%                               | H1             | 80.8%                               | H1             | %                                   |                |                   |                |
| EO Exposure time<br>灭菌暴露阶段 | Min. Load Temp.最低装载温度   | 30.7℃                               | T3             | 31.8℃                               | T1             | ℃                                   |                |                   |                |
|                            | Max. Load Temp.最高装载温度   | 47.4℃                               | T6             | 49.7℃                               | T5             | ℃                                   |                |                   |                |
|                            | Average load Temp.平均装载温度  | 40.0℃                               | N/A            | 42.7℃                               | N/A            | ℃                                   |                |                   |                |
|                            | Min. Load Humidity 最低装载湿度                                       | 71.1%                               | H3             | 42.7%                               | H3             | %                                   |                |                   |                |
|                            | Max. Load Humidity 最高装载湿度                                       | 90.1%                               | H1             | 85.8%                               | H1             | %                                   |                |                   |                |
| Aeration 解析                | Min. Temp.最低温度  | 35.9℃                               | T3             | 39.3℃                               | T1             | ℃                                   |                |                   |                |
|                            | Max. Temp.最高温度  | 47.6℃                               | T6             | 49.6℃                               | T5             | ℃                                   |                |                   |                |

备注 Note: N/A-不存在

结论 Conclusion:

通过全周期循环确认, 观察到的过程参数及装载条件与半周期相似, 物理性参数与半周期相似。由此过程再现性得以证实, 故而, 灭菌循环参数的可靠性可以确认, 产品无菌保证水平 SAL  $10^{-6}$  得到确认。

The cycle parameters and load conditions and load physical profile of the full nominal cycle qualification run similar as half cycle qualification runs observed, hence process reproducibility has been demonstrated. Therefore, the reliable performance of the sterilization cycle for the EO processing group is confirmed and the product SAL  $10^{-6}$  is verified.

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

### 3.8.3. 加药及清洗阶段的压力变化 Pressure rise during gas injection and washing phase

Table- 2 压力变化数据 Pressure rise profile (补充信息，完成后删除此内容)

| Phase 阶段                             | 1 <sup>st</sup> half cycle<br>半周期 1 | 2 <sup>nd</sup> half cycle<br>半周期 2 | 3 <sup>rd</sup> half cycle<br>半周期 3 | Full cycle<br>全周期 |
|--------------------------------------|-------------------------------------|-------------------------------------|-------------------------------------|-------------------|
| EO pressure change (kpa)<br>加药后的压力变化 | 28.5                                | 29                                  |                                     |                   |
| EO charge (min)<br>加药时间              | 11                                  | 7                                   |                                     |                   |
| Rate(kpa/min)速率                      | 2.6                                 | 4.1                                 |                                     |                   |
| Final pressure (kpa)<br>最终压力         | -37.7                               | -37.5                               |                                     |                   |
| Mass of EO used (kg)<br>使用的 EO 量     | 4                                   | 4                                   |                                     |                   |
| EO concentration (mg/L)<br>EO 浓度     | 466.5                               | 474.7                               |                                     |                   |

### 6.5 Method of determining EO gas concentration based on physical laws of perfect gas behavior

#### 6.5.1 100 % EO sterilant

When the sterilant gas injected into the sterilizer chamber is 100 % EO, the mean EO gas density (concentration) within the chamber may be determined by using the relationship shown in Equation (5) and as shown in Example 1 as follows:

$$C_{eo} = \frac{M_{eo(g/mole)} P_{eo(atm)}}{R_{(liter-atm/Kelvin-mole)} T_{(Kelvin)}}, \quad (\text{equation 5})$$

where

$M_{eo}$  = 44.054 g/mole (molecular weight of EO)

$C_{eo}$  = mean EO gas concentration in mg/L

$P_{eo}$  = partial pressure of EO gas injected into the chamber

# Medical Device Manufacturing Co;Ltd

## 医疗器械有限公司

| PHASES<br>阶段             | 1 <sup>st</sup> half cycle 第一次半周期 |                          |                | 2 <sup>nd</sup> half cycle 第二次半周期 |                          |                | 3 <sup>rd</sup> half cycle 第三次半周期 |                          |                | Full cycle 全周期           |                          |                |
|--------------------------|-----------------------------------|--------------------------|----------------|-----------------------------------|--------------------------|----------------|-----------------------------------|--------------------------|----------------|--------------------------|--------------------------|----------------|
|                          | 抽真空时间                             | 压力变化                     | 速率             | 抽真空时间                             | 压力变化                     | 速率             | 抽真空时间                             | 压力变化                     | 速率             | 抽真空时间                    | 压力变化                     | 速率             |
|                          | Time for<br>vacuum (min)          | Pressure<br>change (kpa) | Rate (kpa/min) | Time for<br>vacuum (min)          | Pressure<br>change (kpa) | Rate (kpa/min) | Time for vacuum<br>(min)          | Pressure change<br>(kpa) | Rate (kpa/min) | Time for vacuum<br>(min) | Pressure change<br>(kpa) | Rate (kpa/min) |
| The first vacuum 第一次抽真空  |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| The second vacuum 第二次抽真空 |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| The third vacuum 第三次抽真空  |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| The fourth vacuum 第四次抽真空 |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| The fifth vacuum 第五次抽真空  |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| Air break 1 第一次进气        |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| Air break 2 第二次进气        |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| Air break 3 第三次进气        |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| Air break 4 第四次进气        |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |
| Air break5 第五次进气         |                                   |                          |                |                                   |                          |                |                                   |                          |                |                          |                          |                |

## Medical Device Manufacturing Co;Ltd

### 医疗器械有限公司

对比此次全周期，半周期循环的气体注入阶段及清洗阶段压力的变化，真空度相似，再现性得以证实，由此可以证实物理性能。

Comparing the pressure rise during gas injection and washing phase of this full cycle qualification run and half cycle qualification runs, the vacuum rate are similar, reproducibility has been verified. Therefore the physical performance is qualified

3.8.4. 实施了生物指示剂培养（来自外置挑战装置）。

The biological indicators incubation (from EPCDs) was conducted .

3.8.4.1. 培养结果显示全部 EPCD 都杀灭。

The incubation results showed complete inactivation of the BIs from EPCDs.

3.8.4.2. 阳性对照显示生长，阴性对照显示不生长。

All positives controls showed growth and all negative controls showed no growth.

3.8.4.3. 所有测试结果概述于表 5

Overall test results are summarized in the table # 3 below.

Table 3 - 生物指示剂无菌测试结果 BI Incubation Test Results

| Cycle 循环   | IPCD Results<br>(No. of Positives / Total No. of BI Tested )<br>内置挑战装置 BI 阳性/总数 | EPCD Results<br>(No. of Positives / Total No. of BI Tested )<br>外置挑战装置 BI 阳性/总数 |
|------------|---|---|
| Full cycle | N/A   | 0/9   |

### 3.9. 结论：Conclusions

通过以上常规微生物测试，常规灭菌过程的监控符合接受条件。

Through above routine microbiology testings, current routine sterilization cycle monitoring is qualified.

### 3.10. 产品处置 Disposition of the products

此次全周期循环所采用的产品装载为非销售产品，故无需对产品进行处置。

The load used for this full cycle qualification is made of non-saleable products, there is no products disposal requirements.



#### 4. 纠正措施 **Corrective Actions:**

在实施全周期循环确认过程中未发现偏差，故无需纠正措施。

There is no deviation reported during execution of this full cycle qualification run, hence corrective action is not required.

#### 5. 可交付性 **Deliverables:**

基于物理性能 (PPQ) 和微生物性能 (MPQ) 的确认结果, 现有灭菌柜已经通过性能确认 (PQ), 可用于硅胶引流产品族常规灭菌。

Based on the results of physical performance (PPQ) and microbiological performance qualification (MPQ), existing sterilization chamber is qualified to run the sterilization cycle for current Silicone Drainage Product Family routine sterilization.

#### 6. 附录 **Appendices**

Attachment 1 – 相关人员的培训记录 Training record of relevant staff

Attachment 2 – 仪表校验清单及验证记录 Instruments Calibration review check list and calibration records.

Attachment 3 - 灭菌循环记录 (灭菌报告) Run records (full cycle run report)

Attachment 4 – 装载温湿度分布数据 Load Temperature & humidity data and profiles.

Attachment 5 – 微生物确认测试报告 (生物指示剂) Microbiological Qualification Test Reports (BI incubation results).

Attachment 6 – EO 气体及生物指示剂符合性证书 Certificates of compliance of EO gas and Biological indicator.



医课汇  
公众号  
专业医疗器械资讯平台  
WECHAT OF  
HLONGMED



hlongmed.com  
医疗器械咨询服务  
MEDICAL DEVICE  
CONSULTING  
SERVICES



医课培训平台  
医疗器械任职培训  
WEB TRAINING  
CENTER



医械宝  
医疗器械知识平台  
KNOWLEDG  
ECENTEROF  
MEDICAL DEVICE



MDCPP.COM  
医械云专业平台  
KNOWLEDG  
ECENTEROF MEDICAL  
DEVICE