


Medical Devices Regulation Impact on Resources

Suzie Halliday & Jay Katta

26 July 2016



Impact on Resources


1. Routes of Conformity
2. Certificate Requirements
3. Clinical Evidence
4. Post Market Surveillance
 1. Periodic Safety Update Report
 2. Summary of Safety and Performance
5. Unique Device Identification



Suzie
Halliday




Jay
Katta



Routes of Conformity

Classification & Conformity Assessment – Directive



Classification & Conformity Assessment – Regulatory

Commission Assessment

Competent Authority Assessment

Notified Body Conformity Assessment

Self-Certification

Class III

Class IIb

Risk

Class IIa

Class Im / Is / Ir

Class I

Custom Made

Class III Implants & Class IIb active – delivering medicines

Animal tissues, human tissues, medicinal substances, absorbable

Class IIb Implants

Class IIa – more sampling

Custom Made Class III Implants

Custom Made Devices

Annex XI

Technical Documentation

Annex XIII

PMS / PMCF / Incidents

Name of Person Authorised to make out prescription, Name of Healthcare Institution
& Name of Particular Patient + Meets Requirements of Annex I

Class III Implantable – Custom Made Devices

Annex XI

Technical Documentation

Annex VIII

QMS

Annex X – Part A

Production
Quality Assurance

Name of Person Authorised to make out prescription, Name of Healthcare Institution
& Name of Particular Patient + Meets Requirements of Annex I

CE 0086 CE 0086

Class I Device
(non-sterile / no measuring function / not reusable)

Annex II
Technical
Documentation

Declaration of Conformity (Annex III) & CE Marking (Annex IV)



Class I Device
(sterile / measuring function / reusable)

Annex II
Technical
Documentation


* Only aspects related to
sterility / metrology / reuse
* cleaning, disinfection,
sterilization, maintenance,
functional testing and related
IFU


Annex VIII*
QMS

Annex X – Part A*
Production
Quality Assurance

Declaration of Conformity (Annex III) & CE Marking (Annex IV)







Class IIb Implantable Device

*sutures, staples, dental fillings & braces, tooth crowns, screws, wedges, plates, wires, pins, clips & connectors

Annex VIII

Technical
Documentation

Annex IX

Type Examination

Annex VIII

QMS

Annex X – Part A

Production
Quality Assurance

Annex X – Part B

Product
Verification

Declaration of Conformity (Annex III) & CE Marking (Annex IV)

Class III Device

(including those with medicinal substances, human tissues or animal tissues)

Annex VIII

Technical
Documentation

Annex IX

Type Examination

Annex VIII

QMS

Annex X – Part A

Production
Quality Assurance


Annex X – Part B

Product
Verification


Consultation – 2001/83/EC, EC/726/2004, 2004/23/EC, EU/722/2012


Declaration of Conformity (Annex III) & CE Marking (Annex IV)

Class III Implantable Device & Class IIb Active Devices intended to administer medicinal products* (including those with medicinal substances, human tissues or animal tissues)



Annex VIII – Clause 6 / Annex IX – Clause 6





Certificate Requirements

CE Certificates – Annex XII

bsi.

Quality Management System

CE XXXXXX

Manufacturer: Name + Single Registration Number

Address: Number Street
Town, County
Country, Postal Code

EU Authorised Representative: Name and Address

Scope:

Class III Implantable	Intended Purpose
Device – Generic Device Group	Intended to administer per IFU
Class IIb Active	
Device – Generic Device Group	Intended to administer medicines
Class III	
Device – Generic Device Group	Intended purpose as per IFU
Class IIb Implantable	
Device – Generic Device Group	Intended purpose as per IFU
Class IIb	
Device – Generic Device Group	Intended purpose as per IFU
Class IIa	
Device – Subcategory	---
Class Is	
Class Im	
Class Ir	
Device – Subcategory	---
Custom made Class III Implantable	
Device	---

On the basis of our examination of the quality system under the requirements of Regulation 2016/xx/EU, Annex VIII Chapter I. The quality system meets the requirements of the Regulation. For the placing on the market of Class III and Class IIb implantable devices an Annex VIII Chapter II certificate is required.

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 0086):

Signature Frank Lee
Frank Lee, EMEA Compliance & Risk Director

First Issued: YYYY/MM/DD **Date:** YYYY/MM/DD **Expiry Date:** YYYY/MM/DD

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of this contract.
Information and Contact: BSI, Elm Park Court, Davy Avenue, Knowlton, Milton Keynes MK3 9PP. Tel: +44 (0)1509 9000.
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

...making excellence a habit.

bsi.

Technical Documentation

CE XXXXXX

Manufacturer: Name + Single Registration Number

Address: Number Street
Town, County
Country, Postal Code

EU Authorised Representative: Name and Address

Scope:

Device	Intended Purpose	Classification	UDI-DI
Name, model, type		Class III Implantable	
		Class III	
		Class IIb Implantable	
	Administer Medicinal Substances	Class IIb Active	

BSI has performed a technical documentation assessment on the above devices in accordance with 2016/xx/EU, Annex VIII Chapter II. The documentation meets the requirements of the Regulation. For marketing of these devices an additional Annex VIII Chapter I certificate is required.

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 0086):

Signature Frank Lee
Frank Lee, EMEA Compliance & Risk Director

First Issued: YYYY/MM/DD **Date:** YYYY/MM/DD **Expiry Date:** YYYY/MM/DD

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Regulation as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of this contract.

Information and Contact: BSI, Elm Park Court, Davy Avenue, Knowlton, Milton Keynes MK3 9PP. Tel: +44 (0)1509 9000.
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

...making excellence a habit.

Annex V:


A new UDI-DI shall be required whenever there is a change that could lead to misidentification of the device and/or ambiguity in its traceability, in particular any change of one of the following UDI database data elements require a new UDI-DI:

- Brand Name or Trade name
- Device version or model
- Labelled as single use
- Packaged sterile
- Need for sterilization before use
- Quantity of devices provided in a package
- Critical warnings or contraindications: e.g. containing latex or DEHP



Clinical Evidence

Scope and Definitions – Article 2 – Clinical Evidence



MedDev 2.7.1 Rev 3 / MedDev 2.7.1 Rev 4 / MDR – Equivalence

Technical

- be of similar design
- used under similar conditions of use
- have similar specifications and properties (e.g. physicochemical properties such as intensity of energy, tensile strength, viscosity, surface characteristics, wavelength, software algorithms, porosity, particle size, nanotechnology, specific mass, atomic inclusions – nitrocarburising, oxidability)
- use similar deployment methods (if relevant)
- have similar principles of operation and critical performance requirements

Biological

- use same materials or substances in contact with the same human tissues or body fluids
- for a similar kind and duration of contact and similar release characteristics of substances
- including degradation products and leachables
- Exceptions can be foreseen for devices in contact with intact skin and minor components; in these cases risk analysis results may allow the use of similar materials taking into account the role and nature of the similar material. Evaluators should consider biological safety (e.g. ISO 10993) as well as other aspects necessary for a comprehensive demonstration of equivalence. A justification explaining the situation should be provided for any difference.

Clinical

- used for the same clinical condition or purpose (including similar severity and stage of disease)
- at the same site in the body
- in a similar population (including age, gender, anatomy, physiology)
- have same kind of user
- not foreseen to deliver significantly different performances
- have similar relevant critical performance according to the expected clinical effect for a specific intended purpose

MedDev 2.7.1 Rev 4 – A12.2.3 – Clinical data from an equivalent device and other products

- **Equivalent devices**

- The notified body should clearly document its assessment of clinical data presented from an equivalent device as part of a clinical evaluation. This should critically review and conclude on the equivalence or not of the device under assessment to the devices presented as equivalent in terms of their technical, biological and clinical characteristics. The relevance of each dataset from an equivalent device should be clearly evident and assessed by the notified body.
- The notified body should also **assess and document the level of access** to the technical and clinical data from an Equivalent device that the manufacturer has.
- Relevant information may be commercially sensitive / confidential and not available to the manufacturer. The notified body should **challenge the ability of the manufacturer to access information** that are relevant to the demonstration of equivalence.
- **Demonstration of equivalence might be difficult or impossible in case of limited access to the technical documentation of the devices.**



MedDev 2.7.1 Rev 4 – A12.2.3 – Clinical data from an equivalent device and other products



- **Other products**
- For hazard identification and when assessing the benefit/risk profile of the device, the notified body should consider **current knowledge / the state of the art**.
- The notified body should **assess the appropriateness** of the use of data from **benchmark devices, other devices, and medical alternatives**.

Clinical Evaluation and Post-Market Clinical Follow-up – Annex XIII – Part A: Clinical Evaluation

- These characteristics shall be similar to such an extent that there would be no clinically significant difference in the clinical performance and safety of the device.
- Considerations of equivalence must always be based on proper scientific justification.
- Manufacturers must be able to clearly demonstrate that they have sufficient levels of **access** to the data on devices to which they are claiming equivalence in order to justify that claimed equivalence.

Clinical Evaluation and Investigation – Article 49 – Clinical Evaluation

- In the case of implantable devices **and** devices falling within class III, clinical investigations shall be performed except if:
 - the device has been designed by modifications of a device already marketed by the **same manufacturer**
 - the modified device has been demonstrated to be equivalent and this has been endorsed by the Notified Body (Annex XIII)and
 - the clinical evaluation is sufficient to demonstrate conformity with the relevant safety and performance requirements.
- In this case the notified body shall check that the PMCF plan is appropriate and includes post market studies to demonstrate the safety and performance of the device.
- Clinical investigations need not be performed in the following cases – **sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips or connectors** for which the clinical evaluation is based on sufficient clinical data and is in compliance with the relevant product-specific **common specification**, where such a common specification is available

Clinical Evaluation and Investigation – Article 49 – Clinical Evaluation

- A manufacturer of a device demonstrated to be equivalent to an already marketed device not manufactured by him, may not need to perform a clinical investigation provided that the following conditions are fulfilled in addition to what is required in the paragraph above:
 - the two manufacturers have a **contract in place** that explicitly allows the manufacturer of the second device **full access to the technical documentation** on an **on going basis**,
and
 - the original clinical evaluation has been performed in compliance with the requirements of this Regulation,
and
 - the manufacturer of the second device provides clear **evidence** thereof to the notified body.

Clinical Evaluation and Investigation – Article 49 – Clinical Evaluation

- The requirement to perform clinical investigations shall not apply to implantable devices and devices falling into class III:
 - a) which have been lawfully placed on the market or put into service in accordance with Directive 90/385/EEC or Directive 93/42/EEC and for which the clinical evaluation
 - is based on **sufficient clinical data**
 - and
 - is in compliance with the relevant product-specific **common specification** for the clinical evaluation of that kind of device, **where such a common specification is available**;

or

b)




and is in compliance with common specification for the clinical evaluation, if available

Clinical Evaluation and Investigation – Article 49 – Clinical Evaluation

- Except for class III and implantable devices, where demonstration of conformity with general safety and performance requirements based on clinical data is not deemed appropriate, adequate justification for any such exception shall be given based on the results of the manufacturer's risk management and on consideration of the specifics of the interaction between the device and the human body, the clinical performances intended and the claims of the manufacturer.
- The adequacy of demonstration of conformity with the general safety and performance requirements based on the results of non-clinical testing methods alone, including performance evaluation, bench testing and pre-clinical evaluation, has to be duly substantiated in the technical documentation referred to in Annex II.






Post Market Surveillance

- SSCP
- PSUR

Identification and Traceability of Devices – Article 27 – European Databank



Post-market surveillance, vigilance and market surveillance –

Article 60 C – Periodic Safety Update Report

- Per device and where relevant per category or group of devices, manufacturers of devices in class IIa, IIb and III shall prepare a **periodic safety update report** summarising the results and conclusions of the analyses of the gathered post-market surveillance data according to Annex IIa together with a rationale and description of any preventive and corrective actions taken.



- Manufacturers of class IIb and III devices shall update the report at least annually.
- Manufacturers of class IIa devices shall update the report when necessary and at least every two years.
- Manufacturers of devices in class III or implantable devices shall submit reports by means of the electronic system to the notified body. The notified body shall review the report and add its evaluation to the database with details of any action taken. Such reports and the notified body evaluation shall be available to competent authorities through the electronic system.

Throughout lifetime:

- Conclusions of the benefit risk determination
- Main findings of PMCF
- Volume of Sales
- Estimate of the Population that use the device
- Where practicable usage frequency of the device

Identification and Traceability of Devices


– Article 26 – Summary of Safety and Clinical Performance

- In the case of devices classified as class III and implantable devices, the manufacturer shall draw up a summary of safety and clinical performance.
- It shall be written in a way that is clear to the intended user and, if relevant, to the patient and shall be available to the public via EUDAMED.
- The draft of this summary shall be submitted to the notified body and shall be validated by that body. After validation the notified body shall upload this summary report to Eudamed. The manufacturer shall mention on the label or instructions for use where the summary report is available.

Article 49 – Clinical Evaluation

For devices classified as class III and implantable devices, the PMCF report and, if indicated, the summary of safety and clinical performance shall be updated at least annually with these data.

- Manufacturer + SRN
- Device + UDI
- Intended Purpose, Indications, Contra-indications
- Description, previous variant(s), differences, accessories, other products intended to be used in combination
- Possible diagnostic or therapeutic alternatives
- Harmonised Standards / Common Specifications
- Summary of the Clinical Evaluation Report + PMCF
- Suggested profile and training for users
- Information on residual risks, undesirable effects, warnings & precautions



UDI

Article 24 – Unique Device Identification

4. The UDI carrier shall be placed on the label of the device and on all higher levels of packaging. Higher levels of packaging do not include shipping containers.




Article 24 – UDI



*not commercially available on its own

bsi.



*unique @ all levels of packaging

*if significant space constraint on the unit of use package the UDI may be placed on the next higher package level

Article 24 – UDI

4. The UDI carrier shall be placed on the label of the device and on all higher levels of packaging. Higher levels of packaging do not include shipping containers.
- The UDI shall be used for reporting serious incidents and field safety corrective actions in accordance with Article 61.
- The Basic UDI device identifier ('Basic UDI-DI' as defined in Annex V Part C) of the device shall appear on the EU declaration of conformity referred to in Article 17.
- The manufacturer shall keep up-to-date a list of all applied UDI as part of the technical documentation referred to in Annex II.

Report Form Manufacturer's Incident Report Medical Devices Vigilance System (MEDDEV 2.12/1 rev 7)	
Import XML File save	
fill with test data Initial fill with test data In F fill with test data Follow up fill with test data Final new case, keep base data	
Version 3.24 en 2012-12-04	
1 Administrative information	
Recipient (Name of MCA) Address of National Competent Authority	Stamp box
Date of this report	
Reference number assigned by the manufacturer	
Reference number assigned by MCA	
Type of report	
☐ Initial report ☐ Follow-up report ☐ Combined initial and final report ☐ Final report	
Does the incident represent a serious public health threat?	
☐ yes ☐ no	
Classification of incident	
☐ Death ☐ Unanticipated Serious Deterioration in State of Health ☐ All other reportable incident	
Identify to whom the MCA's this report was also sent	
2 Information on the submitter of the report	
Status of submitter	
☐ Manufacturer ☐ Authorized Representative within EEA and Switzerland and Turkey ☐ Others (identify the role)	

Nomenclature system (preferable GMDN)	Nomenclature code
GMDN	
Nomenclature text	
Commercial name/brand name / make	
Model number	Catalogue number
Serial number(s) (if applicable)	Lot/batch number(s) (if applicable)
Software version number (if applicable)	

Numeric description
of device

Article 24 – UDI

4. The UDI carrier shall be placed on the label of the device and on all higher levels of packaging. Higher levels of packaging do not include shipping containers.
- The UDI shall be used for reporting serious incidents and field safety corrective actions in accordance with Article 61.
- The Basic UDI device identifier ('Basic UDI-DI' as defined in Annex V Part C) of the device shall appear on the EU declaration of conformity referred to in Article 17.
- The manufacturer shall keep up-to-date a list of all applied UDI as part of the technical documentation referred to in Annex II.

Declaration of Conformity

Manufacturer: Name, registered trade name or registered trade mark _____

Address: Address of their registered place of business
Where they can be contacted and their location be established _____

EU Authorised Representative: Name and Address _____

Devices:

- Product or trade name, product code, catalogue number or other unambiguous reference allowing identification and traceability of the device that is covered (it may include a photograph, where appropriate) _____
- UDI device identifier _____
- Risk class of the device in accordance with Annex VII _____

References to the relevant harmonised standards / common technical specifications _____

Where applicable, additional information _____

Notified Body: Where applicable, name and identification number
Description of the conformity assessment procedure performed
Identification of the certificate(s) issued _____

A statement that the declaration of conformity is issued under the responsibility of the manufacturer.

A statement that the device is in conformity with this Regulation and, if applicable, with other relevant Union legislation that make provision for the issuing of a declaration of conformity.

Name and function of the person who signs _____

Signature _____ **Indication for and on behalf of whom he/she signs** _____

Date ____ **Place and date of issue** _____

Article 24 – UDI

4. The UDI carrier shall be placed on the label of the device and on all higher levels of packaging. Higher levels of packaging do not include shipping containers.
- The UDI shall be used for reporting serious incidents and field safety corrective actions in accordance with Article 61.
- The Basic UDI device identifier ('Basic UDI-DI' as defined in Annex V Part C) of the device shall appear on the EU declaration of conformity referred to in Article 17.
- The manufacturer shall keep up-to-date a list of all applied UDI as part of the technical documentation referred to in Annex II.

The technical documentation and, if applicable, the summary thereof to be drawn up by the manufacturer shall be presented in a clear, organised, readily searchable and unequivocal way and shall include:

1. DEVICE DESCRIPTION, SPECIFICATION, VARIANTS & ACCESSORIES
 - Device description and specification + **UDI**
 - Reference to previous / similar generations of the device
2. INFORMATION SUPPLIED BY THE MANUFACTURER
3. DESIGN AND MANUFACTURING INFORMATION
4. **GENERAL SAFETY AND PERFORMANCE REQUIREMENTS**
5. RISK/BENEFIT ANALYSIS AND RISK MANAGEMENT
6. PRODUCT VERIFICATION AND VALIDATION
 - Pre-clinical and clinical data
 - Additional information in specific cases

Annex I – Safety & Performance Requirements

- **SPR# 19.2. Information on the label**
- The label shall bear the following particulars:
- Many new requirements ...
- (h) the unique device identification (UDI) carrier according to Article 24 and Annex V Part C.
- (fa) SPR #7.4.5 – list of carcinogenic, mutagenic, toxic to reproduction, having endocrine disrupting properties >0.1% weight by weight



Article 24 – UDI

- 5. Economic operators** shall store and keep, preferably by electronic means, the UDI of the devices which they have supplied or they have been supplied with, if they belong to:
- **class III implantable devices;**
 - the devices, categories or groups of devices determined by a measure referred to in point (a) of paragraph 7.
- **Health institutions** shall store and keep preferably by electronic means the UDI of the devices which they have supplied or they have been supplied with if they belong to **class III implantable devices**.
- For devices **other than class III implantable devices**, Member States shall **encourage, and may require, health institutions** to store and keep, preferably by electronic means, the UDI of the devices which they have been supplied with.
- Member States shall **encourage, and may require, health care professionals** to store and keep preferably by electronic means, the UDI of the devices which they have been supplied with.

Supply Chain



Raw Materials




Warehouse



✓ Distribution



Disposal



✓ Manufacture
EU AR



✓ Import



✓ Health
Institution



✓ Healthcare
Professional

Patient



Article 16 – Implant Card

The manufacturer of an implantable device shall provide together with the device the following:

- ❑ device name
- ❑ serial number
- ❑ batch code or lot number
- ❑ **Unique Device Identification**
- ❑ device model
- ❑ manufacturer name, address and URL of the website
- ❑ any warnings, precautions or measures to be taken by the patient or a healthcare professional with regard to reciprocal interference with reasonably foreseeable external influences, medical examinations or environmental conditions;
- ❑ any information about the expected lifetime of the device and any necessary follow-up;
- ❑ any other information to assure a safe use of the device by the patient
- ❑ including the information in Annex I, Section 19.3 – Instructions for Use

Questions & Answers

1. Routes of Conformity
2. Certificate Requirements
3. Clinical Evidence
4. Post Market Surveillance
 1. Periodic Safety Update Report
 2. Summary of Safety and Performance
5. Unique Device Identification



医课汇
公众号
专业医疗器械资讯平台
WECHAT OF
HLONGMED



hlongmed.com
医疗器械咨询服务
MEDICAL DEVICE
CONSULTING
SERVICES



医课培训平台
医疗器械任职培训
WEB TRAINING
CENTER



医械宝
医疗器械知识平台
KNOWLEDG
ECENTEROF
MEDICAL DEVICE



MDCPP.COM
医械云专业平台
KNOWLEDG
ECENTEROF MEDICAL
DEVICE