American **National Standard**

ANSI/AAMI/ISO 10993-5:1999

Biological evaluation of medical devices— Part 5: Tests for in vitro cytotoxicity



American National Standard

ANSI/AAMI/ISO 10993-5:1999 (Revision of ANSI/AAMI/ISO 10993-5:1993)

Biological evaluation of medical devices—Part 5: Tests for *in vitro* cytotoxicity

Approved 17 June 1999 by Association for the Advancement of Medical Instrumentation

Approved 10 March 2000 by American National Standards Institute, Inc.

Abstract: This part of ISO 10993 describes test methods to assess the *in vitro* cytotoxicity of medical devices.

Keywords: medical devices, biological evaluation, cytotoxicity

Association for the Advancement of Medical Instrumentation 1110 N Glebe Rd, Ste 220, Arlington, VA 22201-5762 Phone 703/525-4890 Fax 703/276-0793 Internet www.aami.org

AAMI Standard

This Association for the Advancement of Medical Instrumentation (AAMI) standard implies a consensus of those substantially concerned with its scope and provisions. The existence of an AAMI standard does not in any respect preclude anyone, whether they have approved the standard or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standard. AAMI standards are subject to periodic review, and users are cautioned to obtain the latest editions.

CAUTION NOTICE: This AAMI standard may be revised or withdrawn at any time. AAMI procedures require that action be taken to reaffirm, revise, or withdraw this standard no later than 5 years from the

Published by

Association for the Advancement of Medical Instrumentation 1110 N Glebe Road, Suite 220 Arlington, VA 22201-5762

© 2000 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

This publication is subject to copyright claims of ISO, ANSI, and AAMI. No part of this publication may be reproduced or distributed in any form, including an electronic retrieval system, without the prior written permission of AAMI. All requests pertaining to this draft should be submitted to AAMI. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, contact Kurt C. Larrick, Director, Technical Publishing, at AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201. Phone: (703) 525-4890, Ext. 239; Fax: (703) 525-1067.

Printed in the United States of America

ISBN 1-57020-135-8

Contents

	Pa	ge			
Con	nmittee representation	. iv			
Bac	Background of AAMI adoption of ISO 10993-5:1999v				
Fore	Forewordvi				
Introductionvii					
1	Scope	. 1			
2	Normative references	. 1			
3	Terms and definitions1				
4	Sample preparation	. 2			
	4.1 General	. 2			
	4.2 Preparation of liquid extracts of material	. 2			
	4.2.1 Principles of extraction	.2			
	4.2.3 Extraction conditions	. 2			
	4.3 Preparation of material for direct-contact tests	. 3			
5	Cell lines	З			
5		. 0			
6	Culture medium	. 3			
7	Preparation of cell stock culture	. 4			
8	Test procedures	. 4			
	8.1 Number of replicates	. 4			
	8.2 Test on extracts	. 4			
	8.3 Test by indirect contact	.4			
	8.4.1 Agar diffusion test	. 5			
	8.4.2 Filter diffusion test	. 5			
	8.5 Determination of cytotoxicity	. 6			
9	Test report	. 6			
10	Assessment of results	. 7			

Committee representation

Association for the Advancement of Medical Instrumentation

The proposed adoption of ISO 10993-5:1999 as an American National Standard was initiated by the AAMI Biological Evaluation Committee, which also functions as a U.S. Technical Advisory Group to the relevant work in the International Organization for Standardization (ISO). U.S. representatives from the AAMI Cytotoxicity Working Group (U.S. Sub-TAG for ISO/TC 194/WG 5) played an active part in developing the ISO standard

The AAMI Biological Evaluation Committee has the following members:

Cochairmen:	Donald F. Gibbons, PhD Donald E. Marlowe
Members:	James M. Anderson, MD, PhD, Case Western Reserve University Sumner A. Barenberg, PhD, Bernard Technologies Eric R. Claussen, PhD, Cobe Laboratories, Inc. Arthur J. Coury, PhD, Society for Biomaterials Roger Dabbah, PhD, U.S. Pharmacopeial Convention, Inc. Donald F. Gibbons, PhD, 3M Jean A. Goggins, PhD, Medtronic, Inc. Donald E. Marlowe, FDA Center for Devices and Radiological Health Daniel E. McLain, PhD, Becton Dickinson Barry F. Page, Consultant, Garner, NC Harold Stanley, DDS, American Dental Association
Alternates:	Edward Mueller, FDA Center for Devices and Radiological Health Sharon Northup, PhD, U.S. Pharmacopeial Convention, Inc. Patrick J. Parks, PhD, MD, 3M Mel Stratmeyer, PhD, FDA Center for Devices and Radiological Health
The AAMI Cyto	toxicity Working Group has the following members:
Cochair:	Sharon Northup, PhD
Members:	Kim Darnell, CR Bard, Inc. Scott R. Dimond, Nelson Labs Carl Hanks, DDS, University of Michigan Thomas Horbett, PhD, University of Washington Medical Center Katharine Merritt, PhD, FDA Center for Devices and Radiological Health Sharon Northup, PhD, U.S. Pharmacopeial Convention, Inc. Patrick J. Parks, PhD, MD, 3M Robert Przygoda, Ethicon Endo-Surgery, Inc.–J&J Connie Shea, St. Jude Medical, Inc. Harold Stanley, DDS, American Dental Association Nancy Stark, PhD, Clinical Design Group, Inc. Paul J. Upman, PhD, North American Science Associates, Inc. Stuart K. Williams, PhD, University of Arizona Health Science Center Dan Wilson, PhD, Baxter Healthcare Corp. Douglas Woodruff, Meadox Medical, Inc.
Alternates:	Carol Benkendorf, PhD, DABT, Clinical Design Group, Inc. Roger Dabbab, PhD, U.S. Pharmacopeial Convention, Inc.

Alternates: Carol Benkendon, PhD, DABT, Clinical Design Group, Inc. Roger Dabbah, PhD, U.S. Pharmacopeial Convention, Inc. Anthony DeMarinis, BS, MS, CQA, CQM, Davol, Inc.–CR Bard Joel Gorski, PhD, North American Science Associates, Inc. Nirmal Mishra, DVM, PhD, FDA Center for Devices and Radiological Health

NOTE—Participation by federal agency representatives in the development of this standard does not constitute endorsement by the federal government or any of its agencies.

Background of AAMI adoption of ISO 10993-5:1999

Tests for In Vitro Cytotoxicity

As indicated in the foreword to the main body of this document (page vi), the International Organization for Standardization (ISO) is a worldwide federation of national standards bodies. The United States is one of the ISO members that took an active role in the development of this standard.

International Standard ISO 10993-5 was developed by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*, to describe test methods to assess the *in vitro* cytotoxicity of medical devices.

U.S. participation in ISO/TC194 is organized through the U.S. Technical Advisory Group for ISO/TC 194, administered by the Association for the Advancement of Medical Instrumentation on behalf of the American National Standards Institute. The U.S. made a considerable contribution to this international standard.

AAMI encourages its committees to harmonize their work with international standards in the area of biological evaluation of medical devices as much as possible in order to help reduce unnecessary repetition of testing. Upon review of ISO 10993-5 the AAMI Biological Evaluation Committee and the AAMI Cytotoxicty Working Group decided to adopt ISO 10993-5 verbatim as a revision of ANSI/AAMI/ISO 10993-5:1993.

AAMI (and ANSI) have adopted other ISO standards. See the Glossary of Equivalent Standards for a list of ISO standards adopted by AAMI which gives the corresponding U.S. designation and the level of equivalency to the ISO standard.

The concepts incorporated in this standard should not be considered inflexible or static. This standard, like any other, must be reviewed and updated periodically to assimilate progressive technological developments. To remain relevant, it must be modified as technological advances are made and as new data come to light.

Suggestions for improving this standard are invited. Comments and suggested revisions should be sent to Standards Department, AAMI, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-5762.

NOTE—Beginning with the foreword on page vi, this American National Standard is identical to ISO 10993-5:1999.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

International Standard ISO 10993-5 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

This second edition cancels and replaces the first edition (ISO 10993-5:1992), which has been technically revised.

ISO 10993 consists of the following parts, under the general title Biological evaluation of medical devices:

- —Part 1: Evaluation and testing
- -Part 2: Animal welfare requirements
- -Part 3: Tests for genotoxicity, carcinogenicity, and reproductive toxicity
- -Part 4: Selection of tests for interactions with blood
- -Part 5: Tests for cytotoxicity: in vitro methods
- -Part 6: Tests for local effects after implantation
- -Part 7: Ethylene oxide sterilization residuals
- -Part 8: Guidance for reference materials
- -Part 9: Framework for the identification and quantification of potential degradation products
- -Part 10: Tests for irritation and sensitization
- -Part 11: Tests for systemic toxicity
- -Part 12: Sample preparation and reference materials
- -Part 13: Identification and quantification of degradation products from polymers
- -Part 14: Identification and quantification of degradation products from ceramics
- -Part 15: Identification and quantification of degradation products from metals and alloys
- -Part 16: Toxicokinetic study design for degradation products and leachables

Future parts will deal with other relevant aspects of biological testing.

Introduction

Due to the general applicability of *in vitro* cytotoxicity tests and their widespread use in evaluating a large range of medical devices and materials, it is the purpose of this part of ISO 10993, rather than to specify a single test, to define a scheme for testing which requires decisions to be made in a series of steps. This should lead to the selection of the most appropriate test.

Three categories of test are listed: extract test, direct-contact test, indirect-contact test.

The choice of one or more of these categories depends upon the nature of the sample to be evaluated, the potential site of use and the nature of the use.

This choice then determines the details of the preparation of the samples to be tested, the preparation of the cultured cells, and the way in which the cells are exposed to the samples or their extracts.

At the end of the exposure time, the evaluation of the presence and extent of the cytotoxic effect is undertaken. It is the intention of this part of ISO 10993 to leave open the choice of type of evaluation. Such a strategy makes available a battery of tests, which reflects the approach of many groups which advocate *in vitro* biological tests.

The numerous methods used and end-points measured in cytotoxicity determination can be grouped into categories of evaluation type:

- a) assessments of cell damage by morphological means;
- b) measurements of cell damage;
- c) measurements of cell growth;
- d) measurements of specific aspects of cellular metabolism.

There are, therefore, several alternative means of producing results in each of these four categories. The investigator should be aware of the categories of test and into which a particular technique fits, in order that comparisons may be made with other results on similar medical devices or materials, and in order that interlaboratory tests may be conducted.

Biological evaluation of medical devices—Part 5: Tests for *in vitro* cytotoxicity

1 Scope

This part of ISO 10993 describes test methods to assess the in vitro cytotoxicity of medical devices.

These methods specify the incubation of cultured cells either directly or through diffusion

- a) with extracts of a device, and/or
- b) in contact with a device.

These methods are designed to determine the biological response of mammalian cells *in vitro* using appropriate biological parameters.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 10993. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 10993 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 10993-1:1997, Biological evaluation of medical devices—Part 1: Evaluation and testing.

ISO 10993-12:1996, Biological evaluation of medical devices—Part 12: Sample preparation and reference materials.

3 Terms and definitions

For the purposes of this part of ISO 10993, the terms and definitions given in ISO 10993-1 and the following apply.

3.1 negative control material: material which, when tested in accordance with this part of ISO 10993, does not produce a cytotoxic response

NOTE—The purpose of the negative control is to demonstrate background response. For example, high-density polyethylene¹ for synthetic polymers, and aluminum oxide ceramic rods for dental material, have been used as negative controls.

3.2 positive control material: material which, when tested in accordance with this part of ISO 10993, provides a reproducible cytotoxic response

NOTE—The purpose of the positive control is to demonstrate appropriate test system response. For example, an organo-tin stabilized poly(vinylchloride)² has been used as a positive control for solid materials and extracts. Dilutions of phenol, for example, have been used as a positive control for extracts.

3.3 reagent control: extraction vehicle without test material subjected to extraction conditions and test procedures

NOTE—For the purposes of this part of ISO 10993, this definition replaces that given in 3.1 of ISO 10993-12:1996.

¹ High-density polyethylene can be obtained from the U.S. Pharmacopeia (Rockville, Maryland, USA) and Food and Drug Safety Center, Hatano Research Institute (Ochiai 729-5, Hadanoshi, Kanagawa 257, Japan). This information is given for the convenience of the user of this part of ISO 10993 and does not constitute an endorsement by ISO of these products.

² Organo-tin poly(vinylchloride) positive control material is available from SIMS Portex Ltd, Hythe, Kent, CT21 6JL, UK, (product No 499-300-000). The ZDEC and ZDBC polyurethanes are available from Food and Drug Safety Center, Hatano Research Institute (Ochiai 729-5, Hadanoshi, Kanagawa 257, Japan). This information is given for the convenience of the user of this part of ISO 10993 and does not constitute an endorsement by ISO of the product.

3.4 culture vessels: vessels appropriate for cell culture, including glass Petri dishes, plastic culture flasks or plastic multiwells and microtitre plates

NOTE—These vessels may be used interchangeably in these methods provided that they meet the requirements of tissue culture grade and are suitable for use with mammalian cells.

3.5 subconfluency: approximately 80% confluency, i.e. the end of the logarithmic phase of growth

4 Sample preparation

4.1 General

The test shall be performed on either an extract of material; and/or the material itself.

Sample preparation shall be in accordance with ISO 10993-12.

4.2 Preparation of liquid extracts of material

4.2.1 Principles of extraction

Extraction conditions should attempt to simulate or exaggerate the conditions of clinical use so as to determine the potential toxicological hazard, without causing significant changes in the test material such as fusion, melting or alteration of the chemical structure.

NOTE—The concentration of any endogenous or extraneous substances in the extract, and hence the amount of exposure to the test cells, depends on the interfacial area, extraction volume, pH, chemical solubility, diffusion rate, osmolarity, agitation, temperature, time and other factors.

4.2.2 Extraction vehicle

For mammalian cell assays, one or more of the following solvents shall be used. The choice of the extraction vehicle(s) shall be justified:

- a) culture medium with serum;
- b) culture medium without serum;
- c) physiological saline solution;
- d) other suitable solvent.

NOTE—The choice of solvent should reflect the aim of the extraction and consideration shall be given to the use of both a polar and a nonpolar solvent. Suitable solvents include purified water, vegetable oil and dimethylsulfoxide (DMSO). DMSO is known to be cytotoxic in selected assay systems at greater than 0.5% (volume fraction) concentrations.

4.2.3 Extraction conditions

4.2.3.1 The extraction shall be performed in sterile, chemically inert closed containers using aseptic techniques, in general accordance with ISO 10993-12.

4.2.3.2 Recommended extraction conditions are:

- a) not less than 24 h at (37 ± 2) °C;
- b) (72 ± 2) h at (50 ± 2) °C;
- c) (24 ± 2) h at (70 ± 2) °C;
- d) (1 ± 0.2) h at (121 ± 2) °C.

The recommended conditions may be applied according to the device characteristics and specific conditions for use.

Extraction procedures using culture medium with serum can only be used under the conditions specified in 4.2.3.2 a).

4.2.3.3 If the extract is filtered, centrifuged or processed by other methods prior to being applied to the cells, this shall be included in the final report (see clause 9). Any pH adjustment of the extract shall be reported. Manipulation of the extract, such as by pH adjustment, could influence the result.

4.3 Preparation of material for direct-contact tests

4.3.1 Materials that have various shapes, sizes or physical states (i.e., liquid or solid) may be tested without modification in the cytotoxicity assays.

The preferred sample of a solid specimen should have at least one flat surface.

Adjustments shall be made for other shapes and physical states.

4.3.2 Sterility of the test specimen shall be taken into account.

4.3.2.1 Test materials from sterilized devices shall be handled aseptically throughout the test procedure.

4.3.2.2 Test materials from devices which are normally supplied non-sterile but are sterilized before use shall be sterilized by the method recommended by the manufacturer and handled aseptically throughout the test procedure.

The effect of sterilization methods or agents on the device should be considered in defining the preparation of the test material prior to use in the test system.

Test materials from devices not required to be sterile in use shall be used as supplied and handled aseptically throughout the test procedure.

4.3.3 Liquids shall be tested by either

- a) direct deposition; or
- b) deposition onto a biologically inert absorbent matrix.

NOTE—Filter discs have been found to be suitable.

4.3.4 If appropriate, materials classed as superabsorbent shall be soaked with culture medium prior to testing to prevent sorption of the culture medium in the testing vessel.

5 Cell lines

5.1 Established cell lines are preferred and where used shall be obtained from recognized repositories.³

5.2 Where specific sensitivity is required, primary cell cultures, cell lines and organo-typic cultures obtained directly from living tissues shall only be used if reproducibility and accuracy of the response can be demonstrated.

5.3 If a stock culture of a cell line is stored, storage shall be at -80 °C or below in the corresponding culture medium but containing a cryoprotectant, e.g., dimethylsulfoxide or glycerol. Long-term storage (several months up to many years) is only possible at -130 °C or below.

5.4 Only cells free from mycoplasma shall be used for the test. Before use, stock cultures should be tested for the absence of mycoplasma.

6 Culture medium

6.1 The culture medium shall be sterile.

6.2 The culture medium with or without serum shall meet the growth requirements of the selected cell line.

NOTE—Antibiotics may be included in the media provided that they do not adversely affect the assays.

The stability of the culture medium varies with the composition and storage conditions. Media containing serum and glutamine shall be stored at 2 °C to 8 °C for no more than one week. Glutamine-containing media without serum shall be stored at 2 °C to 8 °C for no more than two weeks.

6.3 The culture medium shall be maintained at a pH of between 7.2 and 7.4.

³ For example, recommended cell lines are American Type Culture Collection CCL 1 (NCTC clone 929), CCL 163 (Balb/3T3 clone A31), CCL 171 (MRC-5) and CCL 75 (WI-38), CCL 81 (Vero) and CCL 10 [(BHK-21 (C-13)] and V-79 379A. This information is given for the convenience of the user of this part of ISO 10993 and does not constitute an endorsement by ISO of the products named. Other cell lines may be used if they can be shown to lead to the same or more relevant results.

7 Preparation of cell stock culture

7.1 Using the chosen cell line and culture medium, prepare sufficient cells to complete the test. If the cells are to be grown from cultures taken from storage, remove the cryoprotectant, if present. Subculture the cells at least once before use.

7.2 Remove and resuspend the cells by enzymatic and/or mechanical disaggregation, using a method appropriate for the cell line.

8 Test procedures

8.1 Number of replicates

A minimum of three replicates shall be used for test samples and controls.

8.2 Test on extracts

8.2.1 This test allows both qualitative and quantitative assessment of cytotoxicity.

8.2.2 Pipette an aliquot of the continuously stirred cell suspension into each of a sufficient number of vessels for exposure to the extracts. Distribute the cells evenly over the surface of each vessel by gentle rotation.

8.2.3 Incubate the cultures at (37 ± 2) °C in air with or without 5% (volume fraction) carbon dioxide as appropriate for the buffer system chosen for the culture medium.

The test should be performed on a subconfluent monolayer or on freshly suspended cells.

In the colony-forming assay, only an appropriate low cell density shall be used.

8.2.4 Verify the subconfluency and the morphology of the cultures with a microscope before starting the test.

- 8.2.5 Perform the test on
 - a) the original extract; and
 - b) a dilution series of the extract, using the culture medium as diluent.

If monolayers are used for the test, remove and discard the culture medium from the cultures and add an aliquot of the extract or dilution thereof into each of the vessels.

If suspended cells are used for the test, add the extract or dilution thereof into each of the replicate vessels immediately after preparation of the cell suspension.

8.2.6 When a non-physiological extract is used, e.g., water, the extract shall be tested at the highest physiologically compatible concentration after dilution in culture medium.

NOTE—Concentrated culture medium, e.g., 2x, 5x, is recommended for use in diluting aqueous extracts.

8.2.7 Add known aliquots of the reagent blank and the negative and positive controls into additional replicate vessels.

NOTE—A fresh culture medium control may also be tested, if appropriate.

8.2.8 Incubate the vessels, using the same conditions as described in 8.2.3, for an appropriate interval corresponding to the selected specific assay.

8.2.9 After an incubation period of at least 24 h, determine the cytotoxic effects in accordance with 8.5.

8.3 Test by direct contact

8.3.1 This test allows both qualitative and quantitative assessment of cytotoxicity.

8.3.2 Pipette a known aliquot of the continuously stirred cell suspension into each of a sufficient number of vessels for direct exposure to the test sample. Distribute the cells evenly over the surface of each vessel by gentle horizontal rotation.

8.3.3 Incubate the culture at (37 ± 2) °C in air with or without 5% (volume fraction) carbon dioxide as appropriate for the buffer system chosen for the culture medium until the cultures have grown to subconfluency.

8.3.4 Verify the subconfluency and the morphology of the cultures with a microscope before starting the test.

8.3.5 Remove and discard the culture medium. Then add fresh culture medium to each vessel.

8.3.6 Carefully place individual specimens of the test sample on the cell layer in the center of each replicate vessel. Ensure that the specimen covers approximately one-tenth of the cell layer surface.

Exercise care to prevent unnecessary movement of the specimens, as this could cause physical trauma to the cells, for example patches of dislodged cells.

NOTE—When appropriate, the specimen can be placed in the culture vessel prior to the addition of the cells.

8.3.7 Prepare replicate vessels for both the negative-control and positive-control materials.

8.3.8 Incubate the vessels, under the same conditions as in 8.3.3, for an appropriate interval (a minimum of 24 h) corresponding to the selected specific assay.

8.3.9 Discard the supernatant culture medium and determine the cytotoxic effects in accordance with 8.5.

8.4 Test by indirect contact

8.4.1 Agar diffusion test

8.4.1.1 This test allows a qualitative assessment of cytotoxicity. This assay is not appropriate for leachables that cannot diffuse through the agar layer, or that should react with agar.

8.4.1.2 Pipette a known aliquot of the continuously stirred cell suspension into each of a sufficient number of replicate vessels for the test. Distribute the cells evenly over the surface of each vessel by gentle horizontal rotation.

8.4.1.3 Incubate the cultures at (37 ± 2) °C in air with or without 5% (volume fraction) carbon dioxide as appropriate for the buffer system chosen for the culture medium until the cultures have grown to approximate confluence at the end of the logarithmic phase of the growth curve.

8.4.1.4 Verify the subconfluency and the morphology of the cultures with a microscope before starting the test.

8.4.1.5 Remove and discard the culture medium from the vessel. Then mix fresh culture medium containing serum with melted agar to obtain a final mass concentration of agar of 0.5% to 2%, and pipette an appropriate volume into each vessel. Use only agar that is suitable for the growth of mammalian cells in culture. The agar culture medium mixture should be in a liquid state and at a temperature that is compatible with mammalian cells.

NOTE—Agar is available in various molecular weight ranges and purities.

8.4.1.6 Carefully place replicate specimens of the test sample on the solidified agar layer in each vessel. Ensure that the specimen covers approximately one-tenth of the cell layer surface.

Presoak any absorbent material with the culture medium before placing it on the agar to prevent dehydration of the agar.

8.4.1.7 Prepare replicate vessels with both negative-control and positive-control specimens.

8.4.1.8 Incubate the vessels, using the same conditions as described in 8.4.1.3, for between 24 h and 72 h.

8.4.1.9 Examine the cells to determine cytotoxic effects before and after carefully removing the specimens from the agar.

Use of a vital stain, e.g. neutral red, may aid in the detection of cytotoxicity. The vital stain may be added before or after the incubation with the specimen. If the stain is added before the incubation, protect the cultures from light to prevent cell damage elicited by photoactivation of the stain.

8.4.2 Filter diffusion test

8.4.2.1 This test allows a qualitative assessment of cytotoxicity.

8.4.2.2 Place a surfactant-free filter of pore size 0.45 µm into each vessel, and add a known aliquot of the continuously stirred cell suspension into each of a sufficient number of replicate vessels for the test. Distribute the cells evenly over the surface of each filter by gentle rotation.

8.4.2.3 Incubate the cultures at (37 ± 2) °C in air with or without 5% (volume fraction) carbon dioxide as appropriate for the buffer system chosen for the culture medium until the cultures have grown to approximate confluence at the end of the logarithmic phase of the growth curve.

8.4.2.4 Verify the subconfluency and the morphology of the cultures with a microscope before starting the test.

8.4.2.5 Remove and discard the culture medium from the vessels. Then transfer the filters, cell-side down, onto a layer of solidified agar (see 8.4.1.6).

8.4.2.6 Carefully place the replicate specimens of the test sample on the acellular (top) side of the filter. Retain liquid extracts and freshly mixed compounds in non-reactive rings placed on the filter.

8.4.2.7 Prepare replicate filters with both the negative-control and positive-control specimens.

- **8.4.2.8** ncubate the vessels, using the same conditions described in 8.4.2.3, for $2 h \pm 10 min$.
- 8.4.2.9 Carefully remove the specimens from the filter and carefully separate the filter from the agar surface.
- **8.4.2.10** Determine the cytotoxic effects using an appropriate stain procedure.

8.5 Determination of cytotoxicity

- **8.5.1** Determine cytotoxic effects by either qualitative or quantitative means.
 - a) Qualitative evaluation: examine the cells microscopically, using cytochemical stain if desired. Assess changes in, for example, general morphology, vacuolization, detachment, cell lysis and membrane integrity. The change from normal morphology shall be recorded in the test report descriptively or numerically. A useful way to grade test materials is presented below.

Cytotoxicity Scale	Interpretation
0	Noncytotoxic
1	Mildly cytotoxic
2	Moderately cytotoxic
3	Severely cytotoxic

The method of evaluation and the results of the evaluation shall be included in the test report.

b) Quantitative evaluation: measure the parameter cell death, inhibition of cell growth, cell proliferation or colony formation. The number of cells, amount of protein, release of enzymes, release of vital dye, reduction of vital dye or other measurable parameter may be quantified by objective means. The objective measure and response shall be recorded in the test report.

NOTE—For particular methods of determining cytotoxicity, a zero time or baseline cell culture control may be necessary.

8.5.2 Ensure that care is taken in the choice of evaluation methods, as the test results may be invalid if the test specimen releases substances that interfere with the test system or measurement.

NOTE—Materials that may release formaldehyde can only be reliably tested when cell viability is evaluated.

8.5.3 If there are evident differences in the test result for replicate culture vessels, then the test is either inappropriate or invalid.

8.5.4 If the negative, positive and any other controls (reference, medium, blank, reagent, etc.) do not have the expected response in the test system, then repeat the assay(s).

9 Test report

The test report shall include the following details:

- a) description of the sample;
- b) cell line and justification of the choice;
- c) culture medium;
- d) assay method and rationale;
- e) extraction procedure (if appropriate) and if possible the nature and concentration of the leached substance(s);
- f) negative, positive and other controls;
- g) cell response and other observations; and

h) any other relevant data necessary for the assessment of results.

10 Assessment of results

The overall assessment of the results shall be carried out by a person capable of making informed decisions based on the test data. If the results are inconsistent among replicates or invalid, the tests shall be repeated.